

GAGE APPLIED TECHNOLOGIES, INC.  
1 PROVOST, SUITE 200  
LACHINE, QC CANADA H8S 4H2  
Tel: (514) 633-7447 Fax: (514) 633-0770  
prodinfo@gage-applied.com  
www.gage-applied.com



## CERTIFICATE OF TRACEABLE CALIBRATION

Certificate No.: GMF-04122005-797

Manufacturer: Gage

Model: CS1220 - 4M

Serial No.: A16797

Description: A/D Scope Card

Cal Date: 4/12/2005

Date Placed In Service: \_\_\_\_\_\* Due Date: \_\_\_\_\_\*

Gage Recommended Calibration Interval: 12 months

\*The due date may be established (by the customer) by adding the  
"Recommended Calibration Interval" to the "Date Placed In Service."

### INSTRUMENT CONDITION:

Received: Not Applicable, New Product

Returned: Meets all specifications

Gage certifies the above instrument meets or exceeds published specifications and has been calibrated using standards traceable to the National Institute of Standards and Technology (NIST). This certificate shall not be reproduced, except in full, without the written approval of the calibration facility. Scope card calibrated by manufacturer.

### CALIBRATION PROCEDURE:

Factory Automated Test Software:

- MainTest version: Master V18.0
- TestPC version: Slave V9.0

### CALIBRATION EQUIPMENT USED:

| <u>Mfg/Model Type</u>    | <u>ID #</u> | <u>Cal Date</u> | <u>Due Date</u> |
|--------------------------|-------------|-----------------|-----------------|
| WaveTek 9500/1100        | 40349       | Feb. 23, 2005   | Feb. 23, 2006   |
| WaveTek 9530 Active Head | 42753       | Mar.04, 2005    | Mar.04, 2006    |
| WaveTek 9510 Active Head | 43176       | Mar.04, 2005    | Mar.04, 2006    |

Issued By: 

Production Manager: Ray Lefebvre

Certified By: JS

Date Printed: 04/12/2005